

VISAKHAPATNAM PORT TRUST
MEDICAL DEPARTMENT

TENDER FORM

No:IMED/SO/Shadow less lamp/2016,
Dt: 18-03-2016.
Tender Due Date: **11-04-2016**.

You are requested to tender the sealed cover by furnishing the price against shadow less operating lamp for use in Major Operation Theatre of G.J.Hospital on or before **11-04-2016** at 2.00 Pm at CMO's office/G.J.Hospital/VPT.

Description of the Item	Quantity
<p>“LED” Shadow less operating lamps “SURGICRAFT” (Deluxe Quality) <u>with the following specifications</u></p> <p>A) Intensity > 1,00,000 lux B) Size of light feed > 10cm C) Colour > 3000K D) LED Life > 40,000hrs E) Number of LED > 40 F) Number of domes -2 G) Brightness Control – Digital H) Focus – Adjustable I) Servicing –Servicing Engineer should be available Locally J) Warranty – 2 years</p>	01 No.

For Visakhapatnam Port Trust	All the terms & conditions including those at the overleaf are accepted.
CHIEF MEDICAL OFFICERi/c.	Place: _____ Signature of tenderer with _____ Firm's seal. Dt. _____

Terms & conditions of tender

- 1. The Tenderer should be submitted the sealed cover by furnishing the price** against shadow less operating lamp for use in Major Operation Theatre of G.J.Hospital on or before **11-04-2016** at 2.00 Pm at CMO's office/G.J.Hospital.
2. Quoted price should be free from corrections / erasures. In case there is any unavoidable correction it should be properly attested. If not the quotation will not be considered.
3. The quotations must be in the prescribed form furnished by the port trust.
4. The port Trust reserves the right to reject any or all tenders without assigning any Reasons there of and does not bind itself to accept the lowest quotations.
5. The Port Trust reserves the right to modify the quantity specified in this enquiry

TENDER FOR AMC FOR HOSPITAL EQUIPMENTS
(TO FILLED IN & SIGNED BY THE TENDERER)

To,
The Chief Medical Officer,
Visakhapatnam Port Trust,
Visakhapatnam – 530 035.

No: IMED/SO/Shadow Less Lamp/2016
Dt: 18-03-2016.

To be submitted by 2.00 PM on 11 -04-2016
And to be opened at 3.00 PM on the Same date.

Sir,

Sub: Tender No: _____ for price against
shadow less operating lamp for use in Major Operation Theatre of
G.J.Hospital –Reg;

I/We engage for supply of shadow less operating lamp for use in Major
Operation Theatre of G.J.Hospital upon the terms and conditions as specified
in your limited tender.

SI No	Description	Quantity	Unit Rate in Figure & in words	Total Value in figure	Total value in words.
1.	<p>“LED” Shadow less operating lamp “SURGICRAFT” (Deluxe Quality) <u>with the following Specifications</u></p> <p>A) Intensity > 1,00,000 lux B) Size of light feed > 10cm C) Colour > 3000K D) LED Life > 40,000hrs E) Number of LED > 40 F) Number of domes -2 G) Brightness Control – Digital H) Focus – Adjustable I) Servicing –servicing Engineer should be available Locally J) Warranty – 2 years</p>	01 No			

Encl: Tender conditions duly signed & acceptance/signature of the tenderer

Name & complete Address of Tender & Tenderer M/s.