

**TENDER FOR AMC FOR HOSPITAL EQUIPMENTS
(TO FILLED IN & SIGNED BY THE TENDERER)**

To,
The Chief Medical Officer,
Visakhapatnam Port Trust,
Visakhapatnam – 530 035.

No: IMED/SO/Chemical Analyzer
Dt: 03 -10-2015.

To be submitted by 2.00 PM on 12 -10-2015
And to be opened at 3.00 PM on the Same date.

Sir,

Sub: Tender No: _____ for Procurement of Clinical
Chemistry Analyzer for use in G.J. Hospital. –Reg;

I/We engage to supply 1No. of Clinical Chemistry Analyzer upon the terms and conditions as specified in your limited tender. On buy back system.

SI No	Description	Quantity	Unit Rate in Figure & in words	Total Value in figure/ in words.
1.	<p>Chemical Chemistry Analyzer With the following Specifications</p> <p>1. Analyzer should be open system with triple cuvette system like square cuvette, round cuvette and flow through.</p> <p>2. IAD Ion assisted deposition filter with life long Warranty.</p> <p>3. The Analyzer should be equipped with vacuum based pump for aspiration purpose.</p> <p>4. Local availability of Service Engineer with firms address Proof</p> <p>Less on Buy Back</p>	One	(-)	

Encl: Tender conditions duly signed & acceptance/signature of the Tenderer

Name & complete Address of Tender & Tenderer M/s.