

VISAKHAPATNAM PORT TRUST  
MEDICAL DEPARTMENT

\*\*\*

TENDER FORM

No: IMED/SO/ECGMACHINE/2017,

Dt: 10-09-2017.

Tender Due Date: **09 -10 -2017.**

You are requested to tender the sealed cover by furnishing the price for Procure and supply of 3-Channel ECG Machine for use in X-Ray section of G.J. Hospital, Medical Department on or before 09-10-2017 at 2.00 Pm at CMO's office of G.J .Hospital.

SI NO	Description of the Item	Quantity	Unit cost in figures and words	Total cost in figures and rate
1.	3-Channel ECG Machine Specifications: 1. 3-Channel ECG Recording with 12 lead. 2. Simultaneous acquisition and Arrhythmics Detection in Auto Mode. 3. LCD Display enabling the preview of ECG and preview of patient information entry . 4. Internal Memory which allows storage. 5. Compact light weight. 6. Designed for safe use both on power live as well as battery. 7. Provide Rechargeable battery 8. Removes interference from power live muscle tremor etc., with automatic base live correction. 9. Advanced ECG Analysis and interpretation – the Glasgow ECG interpretation Algorithm. . .	01 No		

For Visakhapatnam Port Trust

CHIEF MEDICAL OFFICER.

All the terms & conditions including those at the overleaf are accepted.

Place:

Signature of tenderer with

Dt.

Firm's seal.

**TENDER FOR 3-CHANNEL ECG MACHINE**  
(TO FILLED IN & SIGNED BY THE TENDERER)

To,  
The Chief Medical Officer,  
Visakhapatnam Port Trust,  
Visakhapatnam – 530 035.

Sir,

Sub: Tender No: \_\_\_\_\_ for Procure  
and supply of 3-Channel ECG Machine for use in X-Ray section of G.J. Hospital  
Medical Department –Reg;

-----

I/We engage to accept the Procure and supply of 3-Channel ECG  
Machine for use in X-Ray section of G.J. Hospital, Medical Department and  
conditions as specified in your limited tender.

SI NO	Description of the Item	Quantity	Unit cost in figures and words	Total cost in figures and rate
1.	3-Channel ECG Machine Specifications: 1. 3-Channel ECG Recording with 12 lead. 2. Simultaneous acajuisction and Arrhythmics Detection in Auto Mode. 3. LCD Display enabling the pre view of ECG and preview of patient information entry . 4. Internal Memory which allows storage. 5. Compact light weight. 6. Designed for safe use both on power live as well as battery. 7. Provide Rechargeable battery 8. Removes interference from power live muscle tremor etc., with automatic base live correction. 9. Advanced ECG Analysis and interpretation – the Glasgow ECG interpretation Algorithm.	01 No		

Encl: Tender conditions duly signed & acceptance/signature of the tenderer

Name & complete Address of Tender & Tenderer M/s.

## Terms & conditions of tender

1. The Tenderer should be submitted the sealed cover by furnishing the price for Procure and supply of 3-Channel ECG Machine for use in X-Ray section of G.J. Hospital, Medical Department in the prescribed Proforma on or before **09-10-2017** at 2.00PM at CMO's Office of G.J.Hospital.
2. Quoted price should be free from corrections / erasures. In case there is any Unavoidable correction, it should be properly attested. If not the quotation will not be considered.
3. The quotations must be in the form furnished by the Port Trust.
4. The Port Trust reserves the right to reject any or all tenders without assigning any reasons there of and does not bind itself to accept the lowest quotations.
5. The Port Trust reserves the right to modify the quantity specified in this enquiry.
6. Your service Engineer shall be available locally.

**Name & complete Address of Tender & Tenderer M/s.**