

VISAKHAPATNAM PORT TRUST
MEDICAL DEPARTMENT

TENDER FORM

No: IMED/SO/Physiotherapy Equipments/ 2021,
Dt: 08-12-2021.

Tender Due Date: **30 -12 -2021.**

You are requested to tender the sealed cover by furnishing the price for the following Physiotherapy Equipments for G.J.Hospital on or before 28 -12-2021 at 2.00 Pm at Dy.CMO's office of G.J .Hospital.

S.No	Description of the Item	Quantity	Cost per one item	Total cost
1.	Inferential Therapy machine	01 No		
2.	Ultra Sound machine	01 No		

Conditions:

- 1. Delivery of the above items as per the specifications at golden jubilee Hospital/ VPT.**
- 2. Payment 100% in 30 days after receipt and acceptance of stores through ECS**
- 3. Bill should be submitted in triplicate**
- 4. GST**

TENDER FOR PHYSIOTHERAPY EQUIPMENTS
(TO FILLED IN & SIGNED BY THE TENDERER)

To,
The Chief Medical Officer,
Visakhapatnam Port Trust,
Visakhapatnam – 530 035.

Sir,

Sub: Tender No: _____ for Procurement of
1no. of Interferential Therapy machine and 1 no. of Ultra Sound Machine,
for use in Physiotherapy unit of G.J.Hospital - Reg;

I/We engage to accept G.J. Hospital and conditions as specified in your limited
tender.

S.No	Description of the Item	Quantity	Cost per one item	Total cost
1.	Inferential Therapy machine	01 No		
2.	Ultra Sound Machine .	01 No		

Encl: Tender conditions duly signed & acceptance/signature of the tenderer

Name & complete Address of Tender & Tenderer M/s.

Technical Specifications

1. Computerized Interferential Therapy machine

Number of Channels	:	2 Channels
Maximum output current	:	0-75m A
Operating Frequency	:	2000Hz & 4000 Hz
Base	:	0.150 Hz
Spectrum	:	0-150 Hz
Modes	:	2EL/4EL/4EL45 Vector/4EL90
Sweep Patterns	:	Triangle/Square/Trapezoidal
Operating Voltage	:	220V AC, 50 Hz
Max Power Consumption	:	50 Watts

List of Accessories for the machine

1. Rubber Electrodes 4 pcs
2. Patient Cables 2 pairs
3. Velcro Straps 3 diff. sizes
4. Main cable
5. User manual

1. Ultrasound Therapy Machine

Ultrasound nominal frequency	:	1 MHz
Ultrasound output	:	12 Acoustic watts Max
Ultrasound power density	:	2.5 Watts/cm ²
Mode of operation	:	Continuous & Pulsed 0 to 20 minutes electronic timer with digital display.
Audio alarm	:	At the end of treatment

Terms & conditions of tender

1. The tenderer should submit the sealed cover by furnishing the price against for Procurement of 1no. of Interferential Therapy machine and 1 no. of Ultra Sound Machine, for use in Physiotherapy unit of G.J.Hospital in the prescribed proforma on or before 30 -12 -2021 at 2.00P.M. at CMO office stores/ G.J. Hospital.
2. Single rate with 1 (one) year Comprehensive warranty should be quoted for the subject item irrespective of their sizes and category and if different rates are quoted , the tender will be summarily rejected.
3. The quantity indicated in the schedule is tentative and may vary as per the requirement.
4. Quoted price should be free from corrections / erasures. In case there is any unavoidable correction, it should be properly attested. If not the quotation will not be considered.
5. The quotations must be in the form furnished by the Port Trust.
6. The Port Trust reserves the right to reject any or all the tenders without assigning any reasons thereof and does not bind itself to accept the lowest quotations.
7. The Port Trust reserves the right to modify the quantity specified in this enquiry.
8. Local Service Engineer should be available to attend any complaints regarding the function of the equipment's and phone no and mail ID should be furnished.

Name & complete Address of Tender & Tenderer M/s.

For Visakhapatnam Port Trust

All the terms & conditions including those at the overleaf are accepted.

Place:

CHIEF MEDICAL OFFICER.

Dt.

Signature of tenderer with
Firm's seal.